



## Accesorios

No Especifica

## Cómo pedir

No Especifica

## Ventajas

El **HI715-11** es un conjunto de estándares de calibración de alta calidad que permite a los usuarios validar el rango medio de amoníaco Checker® HC **HI715**. El **HI715-11** incluye una cubeta sellada para cero ppm de amoníaco y una cubeta teñida a una muestra de amoníaco equivalente a 6,0 +/- 0,2 ppm a 25 °C.

**-Reutilizable con estabilidad de larga vida útil.**

**-Se suministra con certificado de análisis:**

- Código de producto
- Numero de lote
- Fecha de caducidad



## CERTIFICATE OF ANALYSIS

Product Name: CALIBRATION CHECKING SET FOR HI 711  
 Product Code: HI 711-11  
 Lot number: SC0372/14  
 Expiration date: October 2017

Standard cuvette	Lot number	Standard Value @ 25°C (ppm)	Passed
A, HI 711-11	2038	1.00 ± 0.05	✓
B, HI 711-11	2039	1.00 ± 0.05	✓

Hanna Instruments certifies these standards meet the stated tolerance limits as indicated. These should be used for quality control to validate instrument function on the specified Hanna instrument only. Store standards in the storage box upright at room temperature. Do not use standard cuvettes.


Instructions for use:

1. Thoroughly wipe standard cuvette with HI731318 lint free cloth and only handle using top of cuvette. The cuvette wall must be dry and free of smudges.
2. At the C.1 prompt place A, HI 711-11 standard cuvette into the holder of the Checker and press operations button.
3. At the C.2 prompt place B, HI 711-11 standard cuvette into the holder and press the operations button.

Acceptable results must be between 0.95 and 1.05 ppm @ 25°C.

File number: CERT711-11\_SC0372/14

QA manager: Bogdan Munteanu



## Video

No Especifica